CERTIFICAT



CERTIFICATE

No. Z1 14 06 14544 108

Holder of Certificate: Fluke Corporation

6920 Seaway Boulevard Everett, WA 98203-9090

USA

Production Facility(ies):

47951

Certification Mark:



Product: Multimeters

(Handheld Digital Multimeter)

Model(s): 113, 114, 114C, 115, 115C, 116, 116C, 117, 117C

Parameters: Rated Input Voltage: 9VDC (1X9V IEC 6LR61 battery)

Measurement Category: CAT IV 600V: 113

CATIII 600V: 114, 114C, 115, 115C, 116, 116C, 117, 117C

Measuring Current: 10A max (115, 115C, 117, 117C only)

Measuring Voltage: 600V max

Protection Class: Reinforced Insulation
Temperature Ambient: -10°C to +50°C

Tested according to: EN 61010-1:2010

EN 61010-2-033:2012

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: PI1304829-100

Date, 2014-07-03

Page 1 of 1



